



TEST CERTIFICATE: QD12-08-F

Customer:		SO Number:	SO
Order No:		Table Type:	
Customer Ref:		Serial No:	
Date:		Inspector:	

Notation	PARALLELISM					
Method	<p>To determine the deviation in parallelism of a point on the rotary surface and the bottom (seating) surface of the rotary table.</p> <p>Parallelism error is defined as the height difference f_p between 2 planes parallel to the seating surface within which all measured points must be.</p> <p>The rotary table assembly and a high-resolution linear probe and stand are placed on a granite surface plate. The probe is positioned on the top at a diameter position of approximately 0.9 x diameter and zeroed. The stand and probe are not moved but the complete rotary table is slid on the granite and re-positioned where the probe is now at a position approximately 90° from the original on the table top. The table top is rotated 90° in the reverse direction to the base movement. The dial indicator reading is noted and the process repeated at 180°, 270° and 360° positions.</p>					
	Illustration of Test					
References						
Measuring Equipment	Granite Surface Plate					
	High Resolution Indicator					
Measured Results	Angular Position °	0	90	180	270	360
	Deviation f_p (mm)					
	f_p max =	mm				
Notes						

ROTARY PRECISION INSTRUMENTS UK LTD

The Maltings Industrial Estate, Brassmill Lane
Bath BA1 3JL, United Kingdom